

Manual Supplement

Manual Title: 5725A Getting Started
Part Number: 1780175
Print Date: April 2014
Revision/Date:

Supplement Issue: **3**
Issue Date: 7/14
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1 50291

On page 6, under **Relative humidity** add the following:

Pollution Degree: 2

Altitude: 2000 m

Change #2, 58419

On page 6, replace **Safety and EMI/RFI** with the following:

Safety: Meets CAN/CSA C22.2 No. 61010-1-04, UL Std. No. 61010-1 (2nd Edition), and IEC/EN 61010-1:2001

EMI/RFI: Meets IEC/EN 61326:2006 and FCC Rules part 15 subpart B