Manual Supplement

Manual Title:5725A InstructionPrint Date:January 1989Revision/Date:7, 6/02

Supplement Issue:4Issue Date:7/14Page Count:1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	5725A
CD Rev. & Date:	1, 4/14
CD PN:	1780182



Change #1, 50291

On page 1-29, under Relative humidity add the following:

Pollution Degree: 2

Altitude: 2000 m

Change #2

On page 2-3, replace Table 2-1 with the following:

Table 2-1.	Standard	Equipment
	otunidulu	Equipment

Item	Model or Part Number
Amplifier	5725A
Line Power Cord	(See Table 2-2 and Figure 2-1)
Shielded 5700A/5725A Interface Cable	2720557
5725A CD-ROM (contains Instruction Manual)	1780182
Spare 4A, 250V Fuse	216846
Certificate of Calibration	(No Part Number)

Change #3, 58419

On page 1-29, replace Safety and EMI/RFI with the following:

Safety: Meets CAN/CSA C22.2 No. 61010-1-04, UL Std. No. 61010-1 (2nd Edition), and IEC/EN 61010-1:2001 **EMI/RFI:** Meets IEC/EN 61326:2006 and FCC Rules Part 15 subpart B