# INSTRUCTION MANUAL



Serial Number

070-1304-00

Tektronix, Inc. - P.O. Box 500 - Beaverton, Oregon 97005 - Phone: 644-0161 - Cables: Tektronix

.

#### WARRANTY

Tektronix warrants that this product is free from defects in materials and workmanship. The warranty period is one (1) year from the date of shipment. Tektronix will, at its option, repair or replace the product if Tektronix determines it is defective within the warranty period and if it is returned, freight prepaid, to a service center designated by Tektronix.

Tektronix is not obligated to furnish service under this warranty:

- to repair damage resulting from attempts by personnel other than Tektronix representatives to install, repair, or service the product;
- to repair damage resulting from improper use or from connecting the product to incompatible equipment;
- c. If personnel other than Tektronix representatives modify the hardware or software.

There is no implied warranty of fitness for a particular purpose. Tektronix is not liable for consequential damages.

Copyright © 1972 Tektronix, Inc. All rights reserved. Contents of this publication may not be reproduced in any form without the written permission of Tektronix, Inc.

Products of Tektronix, Inc. and its subsidiaries are covered by U.S. and foreign patents and/or pending patents.

TEKTRONIX, TEK, SCOPE-MOBILE, and W are registered trademarks of Tektronix, Inc.

Printed in U.S.A. Specification and price change privileges are reserved.

# TABLE OF CONTENTS

	Page
INSTALLATION PROCEDURE	1
OPERATING INSTRUCTIONS	4
POWERING UP	4
BUILDING A SYSTEM	5
SPECIFICATIONS	6
SYSTEM MAINTENANCE	7
GENERAL	7
TROUBLESHOOTING AIDS	8
TROUBLESHOOTING TECHNIQUES	8
PARTS ORDERING AND REPLACEMENT	10
SERVICE INFORMATION	14
INTERFACE PIN ASSIGNMENTS	15
DETAILED BLOCK DIAGRAM	15
PARTS LOCATION GRID	16
ELECTRICAL PARTS LIST	17
SCHEMATIC	19
MECHANICAL PARTS LIST	20
EXPLODED VIEW DRAWING	

i

# **INSTALLATION PROCEDURE**

## Before you start...

1. Check the rear panel markings. If the factory settings are compatible with the available line voltage and frequency, insert the desired plug-ins. Use the bail to raise the front of the instrument.

## ... go to Operating Instructions...

2. If a change is needed, follow these steps:

#### a. Line Selector Block(s)

Remove the two hold-down screws on the top of the dust cover cabinet and lift the cabinet off. This gives easy access to the Line Selector blocks located on the main circuit board.



#### b. Line Range Taps



#### c. Rear Panel





3. Replace the cabinet.

4. If necessary, change the line cord power plug to match the power source receptacle or use an adapter.

- 5. Plug the cord into the power source.
- 6. Insert the desired plug-ins.

```
7. Use the bail to raise the front of the instrument.
```

## ...go to Operating Instructions...

#### Universal Transformer (SN BO50000-up)

Line Selector	Regulating Ranges				
Block Position	120-Volts Nominal	220-Volts Nominal			
L	90 VAC to 110 VAC	180 VAC to 220 VAC			
M	99 VAC to 121 VAC	198 VAC to 242 VAC			
· H	108 VAC to 132 VAC	216 VAC to 264 VAC			
Line Fuse Data	0.6 A slow-blow	0.3 A med-blow			



TM 501 Power Module with plug-in.

# TM 501 OPERATING INSTRUCTIONS

## INTRODUCTION

#### Description

The TEKTRONIX TM 501 Power Module is a singlecompartment wide mainframe for the TM 500 Series of Modular Instrumentation. It is a basic power source for the many plug-in module members of the TM 500 Series family. Factory options allow customer modification using signal access at the plug-in module/power module interface to provide rear panel inputs and outputs.

#### **Power Source**

This instrument is designed to operate from a power source with its neutral at or near earth (ground) potential with a separate safety-earth conductor. It is not intended for operation from two phases of a multi-phase system, or across the legs of a single-phase, three-wire system.

#### **Power Usage**

The TM 501 can require up to 35 W of power at the upper limits of the high line voltage ranges. Actual power consumption depends on the particular module and operating mode selected.

#### **Operating Temperatures**

The TM 501 is designed to operate in ambient temperatures between  $0^{\circ}$ C and  $+50^{\circ}$ C. Before operating the TM 501 after storage in temperatures within the specified storage range, allow the chassis to return to room ambient before applying power.

## POWERING UP

#### **Plug-in Modules**



Turn the Power Module off before inserting the plugin; otherwise, damage may occur to the plug-in circultry.

#### Module Installation

1. Check the location of the white plastic barriers on the interconnecting jack to insure that their locations match the slots in the edge of the plug-in module's circuit board.

2. Align the plug-in module chassis with the upper and lower guides of the compartment. Push the module in and press firmly to seat the circuit board in the interconnecting jack. (Remove the plug-in module by pulling on the white release latch located in the lower left corner of each module.) 3. Pull the POWER button on the right side of the TM 501. Some plug-in modules have independent power switches, usually labeled OUTPUT, controlling application of mainframe power to the module itself. Push this button to activate the plug-in module.

#### Loading Considerations

The TM 501 can require up to 35 W of power from the line at high-line voltage range settings. Actual power consumed, of course, depends on the particular module selected. This power capability can best be utilized by carefully planning the external loads and the resulting power distributions. Optimum conditions would dissipated as much power as possible in external loads in an ambient temperature around  $+25^{\circ}C$ .

The TM 501 provides the plug-in module with access to a pair of heat-sinked, chassis-mounted transistors, one NPN and one PNP. These Series-Pass transistors allow plug-in modules to operate in power ranges not possible if the power had to be dissipated in the modules themselves.

## BUILDING A SYSTEM

#### **Family Compatibility**

Mechanically, the plug-in modules are very similar to other TEKTRONIX product families. However, they are not **electrically** compatible. Therefore, the TM 501 interface has barriers on the mating connectors between pins 6 and 7 to insure that incompatible plug-ins cannot be inserted. See Figure 1. A compatible module will have a matching slot between pins 6 and 7 of its main circuit board edge connector. This slot and barrier combination is the primary keying assignment.

Another identifier for TM 500-compatible plug-in modules is the white color of the release latch.

#### **Customizing the Interface**

The modularity of this instrumentation system provides for a host of functions to be performed by the plug-in modules. Specific functions are grouped into families or classes, of which there may be several plug-in module members. For instance, some classes are Power Supplies, Signal Sources, Measurement and so forth. Each modular member of a functional family will have a second slot peculiar to its family assignment located in its edge connector. The TM 501 user can "program" the Power Module to accept only members of a certain family by installing a second barrier in the interface connector to match the module's slot location. For extra barriers, order TEKTRONIX Part Number 214-1543-02.

#### **Rear Panel**

The rear sub-panel is punched for BNC and multi-pir connector mountings. Customer- or factory-installed connectors and wiring (see following description of catalog Option 2) could provide external access to the interface for external I/O control. This feature makes the TM 500 Series Modular Instrumentation System very flexible.



Figure 1. Keying assignments for family functions.

**Option 2.** This factory-installed option adds 25-mil squarepin connectors to the rear of the interconnecting jack at all pin locations from pins 14A and B through pins 28A and B. This will keep the interface flexible by making it easy and fast to change customized wiring using prepared wires with squarepin receptacles and long-nose pliers or

tweezers. It also protects the circuit board from damage by repeated soldering and unsoldering of jumper wires. This option also adds a BNC connector and a 50-pin connector to the rear panel. These connectors are **not** pre-wired in order to give a system designer as much flexibility as possible. Instead, prepared jumpers, coax cables, and interconnection jack barriers are included in a kit.

## SPECIFICATIONS

#### **Power Requirements**

Line Voltage Ranges. Universal Transformer. 100, 110, 120, 200, 220, 240 VAC, all within 10%. Range changing for transformer accomplished with quick-change line-selector blocks.

#### Supplies (Unregulated)

Two 25 VAC windings, 500 mA each.

+33.5 V and -33.5 V, 500 mA, maximum.

17.5 VAC and +11.5 V, 1.0 A, maximum, shared in any combination between these two supplies.

#### NOTE

Current and voltage ratings are for main frame maintenance only. In practice, not all available power may be used at once. More detailed information is supplied with kit 040-0652-03 (TM 500-plug-in).

#### **Temperature Range**

Operating:  $0^{\circ}$ C to +50°C. Nonoperating: -40°C to +75°C.

#### Altitude Range

Operating: To 15,000 feet. Nonoperating: To 50,000 feet.

#### Other

Dimensions with Feet and Bail: H 6.0 in., W 3.9 in., L 15.3 in.

Weight without Plug-ins: Approximately 6.0 lb.

#### **Power Cord Conductor Identification**

Conductor	Color	Alternate Color		
Ungrounded (Line)	Brown	Black		
Grounded (Neutral)	Blue	White		
Grounding (Earthing)	Green-Yellow	Green-Yellow		

Line Frequency Ranges. Universal Transformer: 48 Hz to 440 Hz.

**Power Consumption.** Maximum primary power approximately 35 W at high line. Actual power consumption depends on plug-in selection and operating modes.

## TM 501 System Maintenance

## GENERAL

#### Introduction

This section of the manual is meant to support the entire TM 500 Series family of modules with a general coverage of the most commonly-needed service information pertinent to preventive maintenance, troubleshooting, ordering parts, and replacing components and sub-assemblies.

#### **Cabinet Removal**

## WARNING

Dangerous potentials exist at several points throughout the system. When the system must be operated with the cabinet removed, do not touch exposed connections or components. Some transistors have voltages present on their cases. Disconnect power before cleaning the system or replacing parts.

Two screws on the top secure the cabinet to the TM 501 frame. Remove them and lift the cabinet straight up. Do not operate the system with the cabinet removed any longer than necessary for troubleshooting and calibration. Re-install the cabinet to protect the interior from dust and to remove personnel shock hazards.

#### Cleaning



Avoid using chemical cleaning agents which might damage plastic parts. Avoid chemicals containing benzene, toluene, zylene, acetone, or similar solvents.

**Exterior.** Loose dust may be removed with a soft cloth or a dry brush. Water and a mild detergent may be used; however, abrasive cleaners should not be used.

Interior. Cleaning the interior of a unit should precede calibration since the cleaning processes could alter the settings of calibration adjustments. Use low-velocity compressed air to blow off accumulated dust. Hardened dirt can be removed with a soft, dry brush, cotton-tipped swab, or a cloth dampened in a solution of water and mild detergent.

#### **Preventive Maintenance**

Preventive maintenance steps performed on a regular basis will enhance the reliability of the instrumentation system. However, periodic checks of the semiconductors in the absence of a malfunction are not recommended as preventive maintenance measures. See the semiconductor checking information under Troubleshooting Techniques which follow. A convenient time to perform preventive maintenance is just before instrument calibration.

#### Calibration

To insure accurate signal generation and measurement, the performance of individual units comprising the system should be checked periodically. Refer to the Instruction Manual for each unit for complete calibration and verification procedures.

#### **Repackaging for Shipment**

If the Tektronix instrument is to be shipped to a Tektronix Service Center for service or repair, attach a tag showing: owner (with address) and the name of an individual at your firm that can be contacted, complete instrument serial number and a description of the service required.

Save and re-use the package in which your instrument was shipped. If the original packaging is unfit for use or not available, repackage the instrument as follows:

Surround the instrument with polyethylene sheeting to protect the finish of the instrument. Obtain a carton of corrugated cardboard of the correct carton strength and having inside dimensions of no less than six inches more than the instrument dimensions. Cushion the instrument by tightly packing three inches of dunnage or urethane foam between carton and instrument, on all sides. Seal carton with shipping tape or industrial stapler

The carton test strength for your instrument is 200 pounds.

## TROUBLESHOOTING AIDS

#### Introduction

The following is provided to augment information contained elsewhere in this and other TM 500 Series family manuals when troubleshooting becomes necessary.

#### **Circuit Descriptions**

Each manual has a section devoted to explaining circuit operating theory. Used conjointly with the schematics, this can be a powerful analytic tool.

#### Diagrams

Block diagrams and detailed circuit schematics are located on foldout pages in the service section of most of the TM 500 Series family manuals. The schematic diagrams show the component values and assigned circuit reference numbers of each part necessary to the circuit design. Usually the first page of the service sections defines the circuit symbols and reference designators used in that particular instrument. Major circuits are usually identifiable by a series of component numbers. Important waveforms and voltages may be shown within the diagrams or on adjoining aprons. Those portions of the circuits located on circuit boards are enclosed with a blue tint outline.

#### Cam Switch Charts

Cam switches shown on the diagrams are coded in comprehensive charts to locate the cam number of the switch contact in the complete switch assembly, counting from the front, or knob end, toward the rear of the switch. The charts also indicate with a solid dot when each contact is closed.

#### **Circuit Board Illustrations**

Line illustrations showing component locations keyed with a grid scheme for each circuit board are usually placed on the back of a foldout page and sequenced as close as possible to an associated schematic. The GRID LOC column in the Electrical Parts Lists keys each component to the Location illustrations.

#### **Component and Wiring Color Codes**

Colored stripes or dots on electrical components signify electrical values, tolerances, etc., according to EIA standards. Components not color-coded usually have information printed on the body. The wiring coding follows the same EIA standards with the exception of the AC power cord of the Power Modules. It is coded like this:

Black	Line
White	Neutral
Green with a	
Yellow stripe	Safety Earth or Ground

#### Testing Equipment

Generally, a wide-band oscilloscope, a probe, and a multimeter are all that is needed to perform basic waveform and voltage checks for diagnostic purposes. The calibration procedures in the manual for each plug-in module list specific test equipment and the features necessary to adequately check out that particular module.

## TROUBLESHOOTING TECHNIQUES

#### Introduction

This troubleshooting procedure is arranged in an order which checks the simple trouble possibilities before proceeding to extensive troubleshooting.

#### **Control Settings**

Incorrect control settings can indicate a trouble that does not exist. If there is any question about the correct function or operation of any control, see the Operating Instructions section of the manual for the instrument involved.

#### System and Associated Equipment

Before proceeding with troubleshooting the TM 500 Series system, check that the instruments in the system are operating correctly. Check for proper interconnection between the power module and the plug-in module. Check the line voltage at the power source. Check that the signal is properly connected and that the interconnecting cables and the signal source are not defective.

The associated plug-in modules can be checked for proper operation quickly by substituting other like units known to be operating properly. If the trouble persists after substitution, then the power module is probably at fault.

#### Visual Check

Inspect the portion of the system in which the trouble is suspected. Many troubles can be located by visual clues such as unsoldered connections, broken wires, damaged circuit boards, damaged components, etc.

#### **Instrument Calibration**

Check the calibration of the suspected plug-in module or the affected circuit if the trouble is obviously in a certain circuit. The trouble may only be a result of misadjustment or may be corrected by re-calibration. Complete calibration instructions are given in the manual for each instrument in the system.

#### **Circuit Isolation**

Note the symptom. It often identifies the circuit in which the trouble is located. When trouble symptoms appear in more than one circuit, check the affected circuits by making waveform and voltage measurements.

Incorrect operation of all circuits often means trouble in the power supplies. Using a multimeter, check first for correct voltages of the individual regulated supplies according to the plug-in module schematics and calibration procedures. Then check the unregulated supplies of the power module. Defective components elsewhere in the instruments can appear as power supply problems. In these instances, suspected circuits should be disconnected from apparently bad power supplies one at a time to narrow the search.

#### Voltages and Waveforms

Often defective components can be located by using waveform and voltage indications when they appear on the schematic or in the calibration procedures. Such waveforms and voltage labels are typical indications and will vary between instruments. To obtain operating conditions similar to those used to take these readings, refer to the first diagram in the service sections.

#### **Component Checking**

If a component cannot be disconnected from its circuit, then the effects of the associated circuitry must be considered when evaluating the measurement. Each for soldered-in transistors and integrated circuits most components can be lifted at one end from the circuit board.

**Transistors and IC's.** Turn the power switch off before removing or replacing any semiconductor.

A good check of transistor operation is actual performance under operating conditions. A transistor can most effectively be checked by substituting a new component for it (or one which has been checked previously). However, be sure that circuit conditions are not such that a replacement transistor might also be damaged. If substitute transistors are not available, use a dynamic tester. Static-type testers are not recommended, since they do not check operation under simulated operating conditions. A suction-type desoldering tool must be used to remove soldered-in transistors; see component replacement procedure for details.

Integrated circuits can be checked with a voltmeter, test oscilloscope, or by direct substitution. A good understanding of the circuit description is essential to troubleshooting circuits using IC's. Operating waveforms, logic levels, and other operating information for the IC's are given in the circuit description information of the appropriate manual. Use care when checking voltages and waveforms around the IC's so that adjacent leads are not shorted together. A convenient means of clipping a test probe to the 14- and 16-pin in-line IC's is with an integrated-circuit test clip. This device also doubles as an extraction tool.

**Diodes.** Do not use an ohmmeter that has a high internal current. High currents may damage the diode.

A diode may be checked for an open or shorted condition by measuring the resistance between terminals. With an ohmmeter scale having an internal source of between 800 mV and 3 V, the resistance should be very high in one direction and very low when the leads are reversed.

**Resistors.** Check the resistors with an ohmmeter. Resistor tolerances are given in the Electrical Parts List in every manual. Resistors do not normally need to be replaced unless the measured value varies widely from the specified value.

**Capacitors.** A leaky or shorted capacitor can be detected by checking resistance with an ohmmeter on the highest scale. Use an ohmmeter which will not exceed the voltage rating of the capacitor. The resistance reading should be high after initial charge of the capacitor. An open capacitor can best be detected with a capacity meter, or by checking whether it passes AC signals.

## PARTS ORDERING AND REPLACING

#### Ordering

Standard Parts. All electrical and mechanical replacement parts can be obtained through the local TEKTRONIX Field Office or representative. However, many of the standard electronic components can be obtained locally in less time than is required to order them from Tektronix, Inc. Before purchasing or ordering replacement parts, check the Parts Lists for value, tolerance, rating, and description. When selecting replacement parts, it is important to remember that the physical size and shape of the component may affect its performance in an instrument. All replacement parts should be direct replacements unless it is known that a different component will not adversely affect the instrument performance.

**Special Parts.** Some parts are manufactured or selected by Tektronix, Inc., to satisfy particular requirements, or are manufactured for Tektronix, Inc., to our specifications. Most of the mechanical parts used in this system have been manufactured by Tektronix, Inc. Order all special parts directly from the local TEKTRONIX Field Office or representative.

**Ordering Procedure.** When ordering replacement parts from Tektronix, Inc., please include the following minimum information:

- 1. Instrument Type (PS 501, SG 502, DC 501, etc.)
- 2. Instrument Serial Number (For example, B010250)

3. A description of the part (if electrical, include the circuit number)

4. TEKTRONIX part number

Please do not return any instruments or parts before receiving directions from Tektronix, Inc.

A listing of TEKTRONIX Field Offices, Service Centers, and Representatives can be found in the TEKTRONIX Product Catalog and Supplements.

#### Replacing

The exploded view drawings associated with the Mechanical Parts Lists, located to the rear of most manuals,

may be especially helpful when disassembling or reassembling individual components or sub-assemblies.

**Circuit Boards.** If a circuit board is damaged beyond repair the entire assembly including all soldered-on components, can be replaced. Part numbers are given in the mechanical parts list for the completely wired (670 prefix) board (388 prefix).

To remove or replace a board, proceed as follows:

1. Disconnect all leads connected to the board (both soldered lead connections and solderless pin connections).

2. Remove all screws holding the board to the chassis or other mounting surface. Some boards may be held fast by plastic mounting clips around the board edges. For these, push the mounting clips away from the circuit board edges to free the board. Also, remove any knobs, etc., that would prevent the board from being lifted out of the instrument.

3. Lift the circuit board out of the unit. Do not force or bend the board.

4. To replace the board, reverse the order of removal. Use care when replacing pin connectors; if forced into place incorrectly positioned, the pin connectors may be damaged.

**Transistors and IC's.** Transistors and IC's should not be replaced unless they are actually defective. If removed from their sockets during routine maintenance return them to their original sockets. Unnecessary replacement or switching of semiconductor devices may affect the calibration of the instruments. When a transistor is replaced, check the operation of the part of the instrument that may be affected.



Figure 2. Semiconductor device lead configurations found in the TM 500 family.

Replacement semiconductors should be of the original type or a direct replacement. Figure 2 shows the lead configuration of the semiconductors used in this instrument system. When removing soldered-in transistors, use a suction-type de-soldering tool to remove the solder from the holes in the circuit board.

An extracting tool should be used to remove the 14- and 16-pin integrated circuits to prevent damage to the pins. This tool is available from Tektronix, Inc. Order TEKTRONIX Part No. 003-0619-00. If an extracting tool is not available, use care to avoid damaging the pins. Pull slowly and evenly on both ends of the IC. Try to avoid having one end of the IC disengage from the socket before the other end.

To replace one of the power transistors mounted on the Power Module chassis adjacent to the interface circuit board, first unsolder the leads. Then, loosen the nuts which clamp the transistor to the chassis. Remove the defective transistor. When replacing the transistor, use a mica washer on the metal tab to increase heat transfer from the transistor to the chassis. Interconnecting Pins. To replace a pin which is mounted on a circuit board, first disconnect any pin connectors. Then, unsolder the damaged pin and pull it out of the board with a pair of pliers. Be careful not to damage the wiring on the board with too much heat. Ream out the hole in the circuit board with a 0.031-inch drill. Remove the ferrule from the new interconnecting pin and press the new pin into the hole in the circuit board. Position the pin in the same manner as the old pin. If the old pin was bent at an angle to mate with a connector, bend the new pin to match the associated pins.

#### NOTE

A pin replacement kit including necessary tools, instructions, and replacement pins is available from Tektronix, Inc. Order TEKTRONIX Part No. 040-0542-00.

**Cam Switches.** Repair of cam-type switches should be undertaken only by experienced maintenance personnel. Switch alignment and spring tension of the contacts must be carefully maintained for proper operation of the switch. For assistance, contact your local TEKTRONIX Field Office or representative.

#### NOTE

A cam-type switch repair kit including necessary tools, instructions, and replacement contacts is available from Tektronix, Inc. Order TEKTRONIX Part No. 040-0541-00.

The cam-type switches consist of rotating cam drums which are turned by front-panel knobs, and sets of spring-leaf contacts mounted on adjacent circuit boards. The contacts are actuated by lobes on the cams. These switches can be disassembled for inspection, cleaning, repair, or replacement as follows:

1. Remove the screws which hold the metal cover on the switch, and lift the cover off the switch. The switch is now open for inspection or cleaning. 2. To completely remove a switch from the circuit board, first remove any knobs or shaft extensions. Loosen the coupling at the potentiometer at the rear of the switch, and pull the long shaft out of the switch assembly.

3. Remove the screws (from the opposite side of the circuit board) which hold the cam drum to the board.

4. To remove the cam drum from the front support block, remove the retaining ring from the shaft on the front of the switch and slide the cam drum out of the support block. Be careful not to lose the small detent roller.

5. To replace defective switch contacts, follow the instructions given in the switch repair kit.

6. To re-install the switch assembly, reverse the above procedure.

**Pushbutton Switches.** The pushbutton switches are not repairable and should be replaced as a unit if defective. Use a suction-type de-soldering tool to remove solder from the circuit board when removing these switches.

**Incandescent Bulbs.** Most of these light bulbs are mounted on the sub-panel using plastic sleeve stand-offs. Unsolder the lead wires and pull the bulb out of the sleeve from the rear of the sub-panel.

Light-Emitting Diodes. LED's used as indicators are mounted on the sub-panels with plastic sleeve sockets similar to the incandescent bulb mountings or they are soldered directly to a sub-assembly and so mounted that they protrude through holes in the panel. In these cases, the sub-assembly must be exposed and the anode and cathode lead orientations carefully noted before unsoldering the defective LED. See Figure 3 for LED lead identifying information.



Figure 3. Light-emitting diode (LED) lead orientation illustration.

**Power Transformer.** Replace the transformer only with a TEKTRONIX direct replacement transformer. Refer to the exploded view drawing at the rear of the Power Module manuals for disassembly of the rear panel to expose the

power transformer. Refer to the schematic diagram colorcoding information for correct wiring. After replacement, check out the power supply voltages before installing a plug-in module.

,

\*

## **DIAGRAMS AND CIRCUIT BOARD ILLUSTRATIONS**

#### Symbols and Reference Designators

Electrical components shown on the diagrams are in the following units unless noted otherwise:

Capacitors = Values one or greater are in picofarads (pF).

Values less than one are in microfarads ( $\mu$ F).

Resistors = Ohms  $(\Omega)$ .

Graphic symbols and class designation letters are based on ANSI Standard Y32.2-1975.

Logic symbology is based on ANSI Y32.14-1973 in terms of positive logic. Logic symbols depict the logic function performed and may differ from the manufacturer's data.

The overline on a signal name indicates that the signal performs its intended function when it goes to the low state.

Abbreviations are based on ANSI Y1.1-1972.

Other ANSI standards that are used in the preparation of diagrams by Tektronix, Inc. are:

- Y14.15, 1966 Drafting Practices.
- Y14.2, 1973 Line Conventions and Lettering.
  Y10.5, 1968 Letter Symbols for Quantities Used in Electrical Science and Electrical Engineering.

The following prefix letters are used as reference designators to identify components or assemblies on the diagrams.



·..

т Т

· · ·

. .

.

. . .

#### INTERFACE ASSIGNMENTS & BLOCK DIAGRAM

TIM 501





REV C, APR 1979

പ്പ

# POWER MODULE INTERFACE PIN ASSIGNMENTS

FRONT

de-L d B · --ŝ 21 20 19 18 16 3 44 33 12 11 10 ŋ œ ~ φ ល ক 26 24 23 17 ∢ ۲ 28 27 25 ments. Refer to plug-in module manuals for specific No permanent I/O assign-Emitter lead of NPN Series-Pass Emitter lead of PNP Series-Pass Base lead of NPN Series-Pass. Base lead of PNP Series-Pass. +11.5 V common return. assignments. ±33.5 V common return. +11.5 V common return. -33.5 V filtered DC. +33.5 V filtered DC. +11.5 V filtered DC. 17.5 VAC winding. 25 VAC winding. 25 VAC winding.

1

·

.



## PARTS LOCATION GRID

NOTE: COMPONENTS SHOWN WITH DASHED LINES ARE LOCATED ON BACK SIDE OF BOARD.

**REV C, APR 1979** 



 $\langle 1 \rangle$ 

## REPLACEABLE ELECTRICAL PARTS

#### PARTS ORDERING INFORMATION

Replacement parts are available from or through your local Tektronix, Inc. Field Office or representative.

Changes to Tektronix instruments are sometimes made to accommodate improved components as they become available, and to give you the benefit of the latest circuit improvements developed in our engineering department. It is therefore important, when ordering parts, to include the following information in your order: Part number, instrument type or number, serial number, and modification number if applicable.

If a part you have ordered has been replaced with a new or improved part, your local Tektronix, Inc. Field Office or representative will contact you concerning any change in part number.

Change information, if any, is located at the rear of this manual.

#### SPECIAL NOTES AND SYMBOLS

X000	Part first added at this serial number
00X	Part removed after this serial number

#### ITEM NAME

In the Parts List, an Item Name is separated from the description by a colon (:). Because of space limitations, an Item Name may sometimes appear as incomplete. For further Item Name identification, the U.S. Federal Cataloging Handbook H6-1 can be utilized where possible.

#### ABBREVIATIONS

ACTR	ACTUATOR	PLSTC	PLASTIC
ASSY	ASSEMBLY	QTZ	QUARTZ
CAP	CAPACITOR	RECP	RECEPTACLE
CER	CERAMIC	RES	RESISTOR
СКТ	CIRCUIT	RF	RADIO FREQUENCY
COMP	COMPOSITION	SEL	SELECTED
CONN	CONNECTOR	SEMICOND	SEMICONDUCTOR
ELCTLT	ELECTROLYTIC	SENS	SENSITIVE
ELEC	ELECTRICAL	VAR	VARIABLE
INCAND	INCANDESCENT	WW	WIREWOUND
LED	LIGHT EMITTING DIODE	XEMR	TRANSFORMER
NONWIR	NON WIREWOUND	XTAL	CRYSTAL

## CROSS INDEX-MFR. CODE NUMBER TO MANUFACTURER

Mfr. Code	Manufacturer	Address	City, State, Zip
01121	ALLEN-BRADLEY COMPANY	1201 2ND STREET SOUTH	MILWAUKEE, WI 53204
03508	GENERAL ELECTRIC COMPANY, SEMI-CONDUCTOR		
	PRODUCTS DEPARTMENT	ELECTRONICS PARK	SYRACUSE, NY 13201
04713	MOTOROLA, INC., SEMICONDUCTOR PROD. DIV.	5005 E MCDOWELL RD, PO BOX 20923	PHOENIX, AZ 85036
14099	SEMTECH CORP.	652 MITCHELL RD.	NEWBURY PARK, CA 91320
56289	SPRAGUE ELECTRIC CO.		NORTH ADAMS, MA 01247
71400	BUSSMAN MFG., DIVISION OF MCGRAW-		,
	EDISON CO.	2536 W. UNIVERSITY ST.	ST. LOUIS, MO 63107
72982	ERIE TECHNOLOGICAL PRODUCTS, INC.	644 W. 12TH ST.	ERIE, PA 16512
80009	TEKTRONIX, INC.	P O BOX 500	BEAVERTON, OR 97077
91637	DALE ELECTRONICS, INC.	P. O. BOX 609	COLUMBUS, NE 68601
91929	HONEYWELL, INC., MICRO SWITCH DIV.	CHICAGO & SPRING STS.	FREEPORT, IL 61032
95238	CONTINENTAL CONNECTOR CORP.	34-63 56TH ST.	WOODSIDE, NY 11377

N	Ckt No.	Tektronix Part No.	Serial/Mod Eff	el No. Dscont	Name & Description	Mfr Code	Mfr Part Number
	A1 A1	670-2023-00 670-3404-00			CKT BOARD ASSY:INTERFACE CKT BOARD ASSY:INTERFACE (OPTION 2 ONLY)	80009 80009	670-2023-00 670-3404-00
	C2 C10 C12 C22 C30 C32	283-0022-00 283-0004-00 290-0577-00 290-0577-00 283-0002-00 283-0002-00			CAP., FXD, CER D1:0.02UF, 1400VDCAC CAP., FXD, CER D1:0.02UF, +80-20%, 150V CAP., FXD, ELCTLT:2000UF, 50V CAP., FXD, ELCTLT:2000UF, 50V CAP., FXD, CER D1:0.01UF, +80-20%, 500V CAP., FXD, CER D1:0.01UF, +80-20%, 500V	72982 56289 56289 72982	
	C35 C40	290-0578-00 283-0004-00			CAP., FXD, ELCTLT:6000UF, 12V CAP., FXD, CER DI:0.02UF, +80-20%, 150V	56289	68D10429 855-558Z5V0203Z
	CR10 CR30 CR30 CR32 CR32	152-0488-00 152-0198-00 152-0198-02 152-0198-00 152-0198-02	B010100 B040000 B010100 B040000	B039999 B039999	SEMICOND DEVICE:SILICON,200V,1500MA SEMICOND DEVICE:SILICON,200V,3A SEMICOND DEVICE:SILICON,200V,3A SEMICOND DEVICE:SILICON,200V,3A SEMICOND DEVICE:SILICON,200V,3A	04713 03508 14099 03508 14099	3N55 FAMILY 1N5624 3SM2 1N5624 3SM2
	F2 F2	159-0043-00 159-0029-00			FUSE,CARTRIDGE:3AG,0.6A,250V,SLOW-BLOW (FOR 120 VOLT OPERATION) FUSE,CARTRIDGE:3AG,0.3A,250V,SLOW-BLOW (FOR 220 VOLT OPERATION)	71400 71400	MDL 6/10 MDL3/10
	J10	131-1078-00			CONNECTOR, RCPT, :28/56 CONTACT	95238	600-1156Y256DF30
	Q10 Q12	151-0373-00 151-0349-00			TRANSISTOR:SILICON,PNP TRANSISTOR:SILICON,NPN,SEL FROM MJE2801	80009 04713	151-0373-00 SJE924
	R3 R12 R22 R30 R35	308-0704-00 301-0202-00 301-0202-00 302-0102-00 315-0511-00			RES.,FXD,WW:8.8 OHM,5%,5W RES.,FXD,CMPSN:2K OHM,5%,0.50W RES.,FXD,CMPSN:2K OHM,5%,0.50W RES.,FXD,CMPSN:1K OHM,10%,0.50W RES.,FXD,CMPSN:510 OHM,5%,0.25W		EB2025
	S2	260-1222-00			SWITCH, PUSH-PUL: 10A, 250VAC	91929	2DM301
	Tl Tl Tl	120-0790-00 120-0791-00 120-0791-01	B010100 B040000 B070000	B039999 B069999	XFMR, PWR, STPDN: 60HZ XFMR, PWR, STPDN: 50-400HZ XFMR, PWR, STPDN: 50-400HZ	80009	120-0790-00 120-0791-00 120-0791-01

1

.

## REPLACEABLE **MECHANICAL PARTS**

#### PARTS ORDERING INFORMATION

Replacement parts are available from or through your local Tektronix, Inc. Field Office or representative.

Changes to Tektronix instruments are sometimes made to accommodate improved components as they become available, and to give you the benefit of the latest circuit improvements developed in our engineering department. It is therefore important, when ordering parts, to include the following information in your order: Part number, instrument type or number, serial number, and modification number if applicable.

If a part you have ordered has been replaced with a new or improved part, your local Tektronix, Inc. Field Office or representative will contact you concerning any change in part number

Change information, if any, is located at the rear of this manual.

#### SPECIAL NOTES AND SYMBOLS

X000 Part first added at this serial number

onx Part removed after this serial number

#### FIGURE AND INDEX NUMBERS

Items in this section are referenced by figure and index numbers to the illustrations.

**FLCTRN** 

ELCTLT

ELEC

ELEM

EQPT

EXT

FLEX

FLH

FR

FT

FXD

HDL

HEX HEX HD

HLCPS HLEXT

HV IC

١D

IDENT

IMPL B

GSKT

FLTR

FSTNR

FIL

FPI

#### **INDENTATION SYSTEM**

This mechanical parts list is indented to indicate item relationships. Following is an example of the indentation system used in the description column.

12345

Name & Description

Assembly and/or Component Attaching parts for Assembly and/or Component ---\*---

Detail Part of Assembly and/or Component Attaching parts for Detail Part . . . . . . . .

Parts of Detail Part Attaching parts for Parts of Detail Part

Attaching Parts always appear in the same indentation as the item it mounts, while the detail parts are indented to the right. indented items are part of, and included with, the next higher indentation. The separation symbol - - - \* - - - indicates the end of attaching parts.

Attaching parts must be purchased separately, unless otherwise specified.

#### **ITEM NAME**

In the Parts List, an Item Name is separated from the description by a colon (:). Because of space limitations, an Item Name may sometimes appear as incomplete. For further Item Name identification, the U.S. Federal Cataloging Handbook H6-1 can be utilized where possible.

в	INCH
#	NUMBER SIZE
ACTR	ACTUATOR
ADPTR	ADAPTER
ALIGN	ALIGNMENT
AL	ALUMINUM
ASSEM	ASSEMBLED
ASSY	ASSEMBLY
ATTEN	ATTENUATOR
AWG	AMERICAN WIRE GAGE
BD	BOARD
BRKT	BRACKET
BRS	BRASS
BRZ	BRONZE
BSHG	BUSHING
CAB	CABINET
CAP	CAPACITOR
CER	CERAMIC
CHAS	CHASSIS
CKT	CIRCUIT
COMP	COMPOSITION
CONN	CONNECTOR
COV	COVER
CPLG	COUPLING
CRT	CATHODE RAY TUBE
DEG	DEGREE
DWR	DRAWER

ABBREVIATIONS

IN

NIP

ŐD

PL

ΡN

ELECTRICAL ELECTROLYTIC ELEMENT ELECTRICAL PARTS LIST EQUIPMENT EXTERNAL FILLISTER HEAD FLEXIBLE FLAT HEAD FILTER FRAME or FRONT FASTENER FOOT FIXED GASKET HANDLE HEXAGON HEXAGONAL HEAD HEXAGONAL SOCKET HELICAL COMPRESSION HEX SOC HELICAL EXTENSION HIGH VOLTAGE INSIDE DIAMETER **IDENTIFICATION** IMPELLER

FLECTBON

INCH INCAND INCANDESCENT INSUL INSULATOR INTE INTERNAL LAMPHOLDER LPHLDR MACHINE масн MECHANICAL MECH MTG MOUNTING NIPPLE NOT WIRE WOUND ORDER BY DESCRIPTION NON WIRE OBD OUTSIDE DIAMETER OVAL HEAD PHOSPHOR BRONZE ÖVH PH BBZ PLAIN or PLATE PLSTC PLASTIC. PART NUMBER PAN HEAD PNH PWR POWER RECEPTACLE RCPT RESISTOR RES RGD RIGID ALF RELIEF RTNR RETAINER SOCKET HEAD SCH SCOPE OSCILLOSCOPE SCREW SCR

SINGLE END SE SECT SECTION SEMICOND SEMICONDUCTOR SHLD SHIELD SHOULDERED SHLDR SOCKET SKT SL SLFLKG SLIDE SELF-LOCKING SLEEVING SLVG SPR SPRING so SQUARE STAINLESS STEEL SST STEEL STE SWITCH SW TUBE TERMINAL TERM THREAD THD тнк THICK TENSION TNSN TAPPING TPG TRH TRUSS HEAD VOLTAGE VAR VARIABLE WITH W/ WSHR WASHER TRANSFORMER XEMB TRANSISTOR XSTR

#### CROSS INDEX-MFR. CODE NUMBER TO MANUFACTURER

Mfr. Code	Manufacturer	Address	City, State, Zip
000AH	STANDARD PRESSED STEEL CO., UNBRAKO DIV.	8535 DICE ROAD	SANTA FE SPRINGS, CA 90670
00779	AMP, INC.	P O BOX 3608	HARRISBURG, PA 17105
04713	MOTOROLA, INC., SEMICONDUCTOR PROD. DIV.	5005 E MCDOWELL RD, PO BOX 20923	PHOENIX, AZ 85036
07707	USM CORP., USM FASTENER DIV.	510 RIVER RD.	SHELTON, CT 06484
12136	PHILADELPHIA HANDLE COMPANY, INC.	1643 HADDON AVENUE	CAMDEN, NJ 08103
12360	ALBANY PRODUCTS CO., DIV. OF PNEUMO		
	DYNAMICS CORPORATION	145 WOODWARD AVENUE	SOUTH NORWALK, CT 06586
16428	BELDEN CORP.	P. O. BOX 1331	RICHMOND, IN 47374
18680	HIGHLAND MFG. CO., THE DIV. OF BUELL		
	INDUSTRIES, INC.	1240 WOLCOTT STREET	WATERBURY, CT 06720
22526	BERG ELECTRONICS, INC.	YOUK EXPRESSWAY	NEW CUMBERLAND, PA 17070
28520	HEYMAN MFG. CO.	147 N. MICHIGAN AVE.	KENILWORTH, NJ 07033
45722	USM CORP., PARKER-KALON FASTENER DIV.		CAMPBELLSVILLE, KY 42718
70485	ATLANTIC INDIA RUBBER WORKS, INC.	571 W. POLK ST.	CHICAGO, IL 60607
73743	FISCHER SPECIAL MFG. CO.	446 MORGAN ST.	CINCINNATI, OH 45206
75915	LITTELFUSE, INC.	800 E. NORTHWEST HWY	DES PLAINES, IL 60016
78189	ILLINOIS TOOL WORKS, INC.		
	SHAKEPROOF DIVISION	ST. CHARLES ROAD	ELGIN; IL 60120
78471	TILLEY MFG. CO.	900 INDUSTRIAL RD.	SAN CARLOS, CA 94070
80009	TEKTRONIX, INC.	P O BOX 500	BEAVERTON, OR 97077
83385	CENTRAL SCREW CO.	2530 CRESCENT DR.	BROADVIEW, IL 60153
86445	PENN FIBRE AND SPECIALTY CO., INC.	2032 E. WESTMORELAND ST.	PHILADELPHIA, PA 19134
86928	SEASTROM MFG. COMPANY, INC.	701 SONORA AVENUE	GLENDALE, CA 91201
91929	HONEYWELL, INC., MICRO SWITCH DIV.	CHICAGO & SPRING STS.	FREEPORT, IL 61032
93907	CAMCAR SCREW AND MFG. CO.	600 18TH AVE.	ROCKFORD, IL 61101
95238	CONTINENTAL CONNECTOR CORP.	34-63 56TH ST.	WOODSIDE, NY 11377

Replaceable	Mechanical	PartsT	M 501
-------------	------------	--------	-------

Fig. & Index No.		Serial/Model No. Eff Dscont	Qty	12345	Name & Description	Mfr Code	Mfr Part Number
1-1	390-0272-00 390-0272-01		1 1	CABINET, WRAPA CABINET, WRAPA	ROUND:	80009 80009	390-0272-00 390-0272-01
-2	211-0622-00		2	SCREW, MACHINE	(ATTACHING PARTS) 5:6-32 X 0.188,TRH,SST,POZ	12360	OBD
			_	. CABINET ASS			
-3	367-0171-00		2	. HANDLE, CARR		12136	845-R-372-140-37
-4	210-0586-00	B010100 B059999	2		M WA:4-40 X 0.25,STL CD PL	83385	211-041800-00
	210-0783-00	B060000	2		0:0.125 OD X 0.357 L,DOMED HEAD	07707	
~5	210-0958-00			. WASHER, FLAT	:0.115 ID X 0.469 INCH OD, STL	78471	OBD
,	210-0993-00		2	. WSHR, SHOULD	DERED: 0.143" ID X 0.75" OD, BRS	86928	OBD
-6	210-0012-00		2 1		C:INTL,0.375 ID X 0.50" OD STL	78189 80009	1220-02-00-0541C 390-0272-00
-7	390-0272-00 384-1158-00			. CABINET, WRA	CON:POWER SWITCH	80009	384-1158-00
8	376-0127-00			COUPLER, SHAFT		80009	376-0127-00
9	260-1222-00			SWITCH, PUSH-P		91929	2DM301
-10	358-0216-00		1	BUSHING, PLAST	TIC:0.257 ID X 0.412 INCH OD	80009	358-0216-00
-11	333-1530-00		1	PANEL, FRONT:		80009	333-1530-00
~12	211-0022-00		2	SCREW, MACHINE	(ATTACHING PARTS) ::2-56 X 0.188 INCH,PNH STL	83385	OBD
-13	351-0334-00	B010100 B029999	1	GUIDE, PL-IN U	**************************************	80009	351-0334-00
-13	351-0379-01	B030000	1	GUIDE, PL-IN:U		80009	351-0379-01
-14	213-0254-00		n.	SCREW, TPG, TF:	(ATTACHING PARTS) 2-32 X 0.250,100 DEG,FLH	45722	OBD
-15	351-0286-00	B010100 B019999	1	GUIDE.PL-IN U		80009	351-0286-00
	351-0286-01			GUIDE, PL-IN U		80009	351-0286-01
	351-0286-02			GUIDE, PL-IN U		80009	351-0286-02
	351-0286-04	B060000 B082889	1	GUIDE, PL-IN U	INI:LOWER, BLACK NYLON	80009	351-0286-04
	351-0286-07	B082890	1	GUIDE, PL-IN U	INI:LOWER,NYLON (ATTACHING PARTS)	80009	351-0286-07
-16	211-0101-00		1		::4-40 X 0.25" 100 DEG,FLH STL	83385	OBD
	213-0814-00	B082890	1	SCREW, TPG, TR:	4-20,0.25 L,PLASTITE	93907	OBD
-17	348-0187-00		4	FOOT, CABINET:	0.780 X 1.650 INCH LONG (ATTACHING PARTS)	80009	348-0187-00
-18	211-0551-00		1	SCREW, MACHINE	::6-32 X 0.562 INCH, PNH STL	83385	OBD
-19	348-0026-00		2	HINGE BLOCK, S	TA:LEFT	80009	348-0026-00
-20	348-0027-00		2	HINGE BLOCK,S		80009	348-0027-00
-21	348-0303-00		1	STAND, ELEC EQ		80009	348-0303-00
-22	131-1018-00				PLUG-IN GROUND	80009	131-1018-00
	131-1254-01	B020000 B075079X	1	CONTACT, ELEC:	(ATTACHING PARTS)	80009	131-1254-01
	211-0008-00	B010100 B075079X	1	SCREW, MACHINE	:4-40 X 0.25 INCH, PNH STL	83385	
-23	210-0586-00	B010100 B075079X	1	NUT,PL,ASSEM	WA:4-40 X 0.25,STL CD PL	83385	211-041800-00
-24	·······		1	CKT BOARD ASS	SY:INTERFACE(SEE A1 EPL)		
-25	131-0608-00				N:0.365 L X 0.25 PH, BRZ, GOLD PL	22526	47357
-26	131-1078-00			-	CPT,:28/56 CONTACT	95238	600~1156Y256DF30
-27	214-1593-02			-	ZN:CKT BD CONN	80009	214-1593-02
-28	131-1895-00 131-0707-00				CONN:8,22 AWG,1.5 L (,TERM.:22-26 AWG,BRS& CU BE GOLD	80009 22526	131-1895-00 47439
29	352-0166-01		2		PL,EL:8 WIRE BROWN	80009	352-0166-01
<i>c</i> , <i>j</i>	131-1896-00		1		CONN:8,22 AWG,1.5 L	80009	131-1896-00
	352-0166-02		1	HLDR, TERM	f CONN:8 WIRE, RED	80009	352-0166-02
	131-0707-00 352-0166-06		2 1		R, TERM.: 22-26 AWG, BRS& CU BE GOLD PL, EL:8 WIRE BLUE	22526 80009	47439 352-0166-06
-30	917_0000 00		4	CCD TDC THE C	(ATTACHING PARTS) CTG:4-24 X 0.25 INCH,PNH STL	83385	OBD
···· 5   1	213-0088-00					83385	OBD
-31	211-0012-00	B010100 B077419	2	SCREW, MACHINE	2:4-40 X 0.375,PNH STL CD PL	03303	050
	211-0012-00 211-0098-00		2  2	(STANDARD ONL		000AH	

. 4

23

#### Replaceable Mechanical Parts-TM 501

						· · · · · ·			
Fig. &		_							2 and the second
Index	Tektronix	Serial/Mod	el No.				Mfr		
No.	Part No.	Eff D	Dscont	Qty	12345	Name & Description	Code	Mfr Part Number	No. of
1-	211-0012-0	р војојоо :	B078539	2		:4-40 X 0.375, PNH STL CD PL	83385	OBD	
	211-0098-0	0 B078540		2		D:4-40 X 0.375 INCH,STL	000AH	OBD	
20				~	(OPTION 2 ONL				
-32 -33	342-0136-0			2		R:0.812 OD X 0.0025 INCH THK	04713		
-33	210-0071-0		0077/10	2		ISN:0.146 ID X 0.323" OD, STL		4706-05-01-0531	
Ju	210-0586-0		6077419	2	(STANDARD ONL	WA:4-40 X 0.25,STL CD PL	83385	211-041800-00	
	220-0601-0					.:4-40 X 0.25 INCH HEX, BRS	73743	93261-03	
	210-0586-0		B078539	2		WA:4-40 X 0.25,STL CD PL	83385	211-041800-00	
	210-0601-0			2		IC:CAD PLATED	18680	A424	
-35		_		1	TRANSFORMER:(				
				1	INHIGPORTER . (	(ATTACHING PARTS)			
-36	212-0576-0	D B010100	8039999	4	SCREW MACHINE	:10-32 X 1.375", HEX, STL	83385	OBD	
	212-0517-0		0000000	4		:10-32 X 1.750INCH, HEX HD STL	83385		
		0 XB052090	8091899	4	WASHER, NONMET		86445		
	210-0010-0		0071077	4		NT,0.20 ID X0.376" OD,STL		1210-00-00-0541C	
	210-0206-0			1	TERMINAL, LUG:			A373-147-1	
-37	220-0410-0			4	•	WA:10-32 X 0.375 INCH,STL	83385		
-38	166-0226-0			4		1.125 INCHES LONG	80009	166-0226-00	
	100 0000 0				Ind DBV, ELLO.		00007	100 0220 00	
-39	200-0379-0	l		1	COVER, ELEC XF	MR:	80009	200-0379-01	
-40	210-0201-00	)	-	1	TERMINAL, LUG:			A373-157-2	
						(ATTACHING PARTS)			
-41	210-0586-00	)		1	NUT, PL, ASSEM	WA:4-40 X 0.25,STL CD PL	83385	211-041800-00	
-42	352-0076-00	D B010100	B049999	1	FUSEHOLDER:W/	HARDWARE	75915	342012-L	
	352-0362-0	в050000		1		/MOUNTING HARDWARE (ATTACHING PARTS)	75915	345001	
-43	210-0873-0	)		1	WASHER, NONMET	AL:0.5 ID X 0.688 INCH OD, NPRN	70485	OBD	Nerrow C
-44	358-0161-0	0		1	BSHG, STRAIN R	LF:FOR 0.50 INCH HOLE, PLASTIC	28520	SR5P4	
-45		4 B010100	B082459	1	·	· · · · · · · · · · · · · · · · · · ·		KH854-2	
	161-0033-2	5 B082460		1		R:3,18 AWG,125V,84.0 L	16428		
	210-0201-0	XB073290		1	TERMINAL, LÚG:	SE #4		A373-157-2	
	210-0586-00	) XB073290		1	NUT, PL, ASSEM	(ATTACHING PARTS) WA:4-40 X 0.25,STL CD PL	83385	211-041800-00	
-46	333-1560-00	)		1	PANEL, REAR:		80009	333-1560-00	
-47	426-0876-00	) B010100 I	B075979	1		Y:CABINET		426-0876-00	
	426-0876-03			1	FRAME ASSEMBL			426-0876-02	
	334-3379-0	L XB080730		1		MARKED GROUNDSYMBOL		334-3379-01	
	198-2070-00	)		1	WIRE SET, ELEC			198-2070-00	
	131-2065-00	) XB084930		3		SC:18-22 AWG, BRASS TIN PLATED		2-350799-2	





4

.

#### STANDARD ACCESSORIES

Index No.	Tektronix Part No.	Serial/Mo Eff	del No. Dscont	Qty	12345	Name &	Description		Mfr Part Numb
	070-1304-00	0		1.	MANUAL, TECH: I	NSTRUCTION		80009	070-1304-00

.

#### MANUAL CHANGE INFORMATION

At Tektronix, we continually strive to keep up with latest electronic developments by adding circuit and component improvements to our instruments as soon as they are developed and tested.

Sometimes, due to printing and shipping requirements, we can't get these changes immediately into printed manuals. Hence, your manual may contain new change information on following pages.

A single change may affect several sections. Since the change information sheets are carried in the manual until all changes are permanently entered, some duplication may occur. If no such change pages appear following this page, your manual is correct as printed.

#### SERVICE NOTE

Because of the universal parts procurement problem, some electrical parts in your instrument may be different from those described in the Replaceable Electrical Parts List. The parts used will in no way alter or compromise the performance or reliability of this instrument. They are installed when necessary to ensure prompt delivery to the customer. Order replacement parts from the Replaceable Electrical Parts List.

# CALIBRATION TEST EQUIPMENT REPLACEMENT

## **Calibration Test Equipment Chart**

This chart compares TM 500 product performance to that of older Tektronix equipment. Only those characteristics where significant specification differences occur, are listed. In some cases the new instrument may not be a total functional replacement. Additional support instrumentation may be needed or a change in calibration procedure may be necessary.

	Comparison of Main Character						
DM 501 replaces 7D13		107 - Risetime less than					
PG 501 replaces 107	PG 501 - Risetime less than	$3.0 \text{ ns into } 50 \Omega.$					
	3.5 ns into 50 Ω.	108 - 10 V output pulse					
108	PG 501 - 5 V output pulse;	1 ns Risetime					
	3.5 ns Risetime	1 H5 Filocuirie					
PG 502 replaces 107							
	PG 502 - 5 V output	108 - 10 V output					
108	PG 502 - Risetime less than	111 - Risetime 0.5 ns; 30					
111	1 ns: 10 ns	to 250 ns					
	Pretrigger pulse	Pretrigger pulse					
	delay	delay					
PG 508 replaces 114		and in the same of					
	Performance of replacement equipment is the same or						
115	better than equipment being replaced	J.					
2101		106 - Positive and Negative-					
PG 506 replaces 106	PG 506 - Positive-going	going trigger output					
- a out replaced top	trigger output sig-	signal, 50 ns and 1 V;					
	nal at least 1 V;	High Amplitude output,					
	High Amplitude out-	100 V.					
	put, 60 V.						
067-0502-01	PG 506 - Does not have	0502-01 - Comparator output can be alternately					
001-0002 01	chopped feature.	chopped to a refer-					
		ence voltage.					
SG 503 replaces 190,		190B - Amplitude range 40 mV					
190A, 190B	SG 503 - Amplitude range	to 10 V p-p.					
	5 mV to 5.5 V p-p.						
191	SG 503 - Frequency range	0532-01 - Frequency range					
067-0532-01	250 kHz to 250 MHz.	65 MHz to 500 MHz.					
SG 504 replaces							
067-0532-01	SG 504 - Frequency range	0532-01 - Frequency range 65 MHz to 500 MHz.					
00, 0002 01	245 MHz to 1050 MHz.	65 WHZ 10 500 WHZ.					
067-0650-00							
TG 501 replaces 180,		180A - Trigger pulses 1, 10,					
180A	TG 501 - Trigger output- slaved to marker	100 Hz; 1, 10, and					
	output from 5 sec	100 kHz. Multiple					
	through 100 ns. One	time-marks can be					
	time-mark can be	generated simultan-					
	generated at a time.	eously.					
	yenerated at a mile	181 - Multiple time-marks					
181	TO SOL Trigger output-	184 - Separate trigger					
184	TG 501 - Trigger output- slaved to market	pulses of 1 and 0.1					
	output from 5 sec	sec; 10, 1, and 0.1					
	through 100 ns. One	ms; 10 and 1 $\mu$ s.					
	time-mark can be						
	generated at a time.						
		2901 - Separate trigger					
2901	TG 501 - Trigger output-	pulses, from 5 sec					
	slaved to marker	to 0.1 µs. Multiple					
	output from 5 sec	time-marks can be					
	through 100 ns. One time-mark can	generated simultan-					
	One time-mark carr	eously.					
	be generated at						

NOTE: All TM 500 generator outputs are short-proof. All TM 500 plug-in instruments require TM 500-Series Power Module. **REV B, JUN 1978** 

.

.