

terial suspension. As indicated, a linear relationship is obtained between the galvanometer reading and the turbidity of the solution over a range of galvanometer readings from 10 to 60.

A 32 candlepower, 6-volt lamp furnished the light source. Current for the lamp was supplied by a constant current voltage regulator. A galvanometer with a sensitivity of 1 mm deflection per 0.125 microamperes was used to record the current generated by the photonic cell. Round bottom 100×13 mm soft glass or pyrex test-tubes were used to hold the turbid solutions.

In checking several hundred of the above test-tubes it was found that only 1 in 10 gave a deviation in reading of over ± 1.0 per cent. with a constant suspension.

RAYMOND L. LIBBY

LEDERLE LABORATORIES, INC.,
PEARL RIVER, N. Y.

AN IMPROVED WATER BOTTLE FOR SMALL "CAGED" ANIMALS

INVESTIGATORS working with albino rats are constantly aware of the inefficiency of the various types of water bottles necessary for large stock cages. Containers which have an opening large enough for the animal to enter its head likewise serve as receptacles for excreta. An improvement over this type of container is the inverted bottle carrying a straight glass tube. Yet, it, too, offers disadvantages, since proper functioning depends on an equal pressure both inside and out. Thus, whenever from one third to one half of the contents are consumed, a vigorous regurgitation is initiated and continues until the container is empty. As a result of such action, the animals may go several hours without water or other liquids. The cages become and remain wet, which gives rise to obnoxious odors and contamination. This problem is particularly evident when the attendants find it impossible to keep a close check on the water bottles.

To meet the problem of inefficient water bottles for rats, an adaptable unit has been devised. The apparatus required is simple and can be put together from materials available in the laboratory. The apparatus used here has a capacity of approximately 500 cc, a number eleven stopper and 7 mm glass tubing. This improved container, schematically shown in Fig. 1, will not leak or regurgitate when properly adjusted. Tubes *C* and *D* should be of equal length when in place.




FIG. 1

This is best attained after the end *C* has been introduced through the stopper. Place the inside of the stopper along the straight edge of a

laboratory table, thus permitting the ends of the tubes *C* and *D* to rest flat against the top surface of the table. Mark both tubes at an even length from the straight edge, cut and fire polish. Tube *A-B* should be a short close bend, and the length between the center of the bend to the end of *B* from 1 to 1½ inches.

Fill the bottle, put stopper in place and invert. Tube *A-B* will serve as an air inlet. By means of the inverted siphon *C-D* the water will flow through the tube without the aid of force. However, since *C* and *D* are at the same height the water will not flow beyond *D* without additional pressure. This pressure is supplied through the air inlet tube *A-B*, whenever the animal removes the slightest amount of water. The ideal distance between *B* and *C* is ½ inch. In the event that the contents flow too freely, the distance between *B* and *C* should be shortened. On the other hand, should the supply be insufficient the distance between *B* and *C* should be lengthened. The container can be attached to the side of the cage. However, the apparatus is conveniently handled when allowed to rest on a simple rack, as shown in Fig. 1. The rack can be made from two pieces of wood nailed at right angles, and permanently fastened to the side of the cage by screws or bolts. It is necessary to saw a block of wood one quarter inch wide from the center of the rack through *xyz* for the passage of the two tubes. This will allow the bottle to slide in place, thus permitting tube *D* to enter the cage.

Through evaporation, removal of water by the animals or changes in temperature, the equilibrium is maintained. The balance is most delicate and remains constant when properly adjusted.

D. S. THORPE

WASHINGTON HIGH SCHOOL,
ATLANTA, GA.

BOOKS RECEIVED

- Bicentennial Conference of the University of Pennsylvania: GOLDBLATT, HARRY, and others. *Hypertension*. Pp. 46. \$0.50. FIESER, LOUIS F., and others. *Cause and Growth of Cancer*. Pp. 64. \$0.75. KLEIN, HENRY, and others. *Dental Caries*. Pp. 53. \$0.50. PETERS, JOHN P., and others. *Problems of Intestinal Obstruction*. Pp. 56. \$0.50. JEWETT, FRANK B., and ROBERT W. KING. *Engineering Progress and the Social Order*. Pp. 15. \$0.25. University of Pennsylvania Press.
- HUNTRESS, ERNEST H. and SAMUEL P. MULLIKEN. *Identification of Pure Organic Compounds*. Pp. xvii + 691. Wiley. \$7.50.
- Social Security Board. Bureau of Research and Statistics. SAKMANN, MARIANNE, and others. *Bureau Report No. 5; An Outline of Foreign Social Insurance and Assistance Laws*. Pp. 62. \$0.15. OTEY, ELIZABETH L. *Bureau Report No. 6; Cash Benefits under Voluntary Disability Insurance in the United States*. Pp. vi + 117. \$0.15. Superintendent of Documents, Washington.
- The Manuscript; a Guide for its Preparation*. Pp. xii + 75. Illustrated. Wiley. \$1.00.