

Nanosphere™ Size Standards. Certified in billionths of a meter by Duke Scientific

Nanosphere Size Standards are calibrated in billionths of a meter (nanometers) and are available in 22 sizes from 21 to 900nm—all traceable to the National Bureau of Standards. Nanospheres are part of our complete line of spherical particles from 0.02 to 2000 micrometers in diameter. They are used as standards for instrument calibration, quality control, filter checking, and in numerous biotechnology applications. At Duke Scientific—established in 1971—we have the expertise and resources to meet any of your requirements for microspheres and particles. Call us today for information.



1135D San Antonio Road, Palo Alto, CA 94303, Toll Free
(800) 334-3883, in CA (415) 962-1100, Fax (415) 962-0718



Circle No. 1 on Readers' Service Card

Air Pollution and Mortality

In our letter (9 Aug. 1991, p. 606) about the article by Alan J. Krupnick and Paul R. Portney (26 Apr. 1991, p. 522), we pointed out uncertainties in the assignment of effects of mortality due to air pollution on the basis of comparisons among U.S. metropolitan areas (an "ecological" study) by Haluk Özkaynak and George D. Thurston (1). Krupnick and Portney agreed with our assessment, but Thurston and Özkaynak (Letters, 24 Jan., p. 382) defend their point estimates. Our basic argument is that uncertainties about magnitude and causal agent are inherent in all ecological studies of this kind. In particular, one of the key findings of Lipfert *et al.* (2) was that a slightly superior *t* statistic provided no guarantee that one model (or pollutant) gave a statistically significantly better fit than its competitors. Furthermore, the significance of a variable is not a measure of its relative toxicity. As one attempts to assign costs of pollution damage to particular pollutants, it is important to recognize the poor state of our ability to fix blame. A "winner take all" approach to estimating benefits is inappropriate, given the wide disparities in control costs for equally plausible "responsible" pollutants

and the small differences in their performance in alternative regression models.

**Frederick W. Lipfert
Samuel C. Morris**

Brookhaven National Laboratory,
Associated Universities, Inc.,
Upton, NY 11973

REFERENCES

1. H. Özkaynak and G. D. Thurston, *Risk Anal.* 7, 449 (1987).
2. F. W. Lipfert *et al.*, "A statistical study of the macroepidemiology of air pollution and total mortality" (Report 52122, Brookhaven National Laboratory for the U.S. Department of Energy, Upton, NY, 1988).

Corrections and Clarifications

In Mark R. Patterson's report "A mass transfer explanation of metabolic scaling relations in some aquatic invertebrates and algae" (13 Mar., p. 1421), the formula for the Reynolds number two lines above equation 1 should have read, " $Re = \rho UW/\mu$." In table 2 (p. 1422), the mass exponents for the flat plate of laminar and turbulent flow should have been 0.75 and 0.90, respectively. In the abstract, the range of exponents given in line 6 should have been 0.31 to 1.00.

In the ScienceScope article "Emergency rescue for Soviet germ plasm" (3 Apr., p. 19), samples of germ plasm were incorrectly identified as "ascensions." They are "accessions."

The ideal way to measure osmolality.

The biotechnology explosion has expanded the need for measuring the osmolality of solutions. Such measurements are critical in many areas of research. The most current and accurate means of measuring osmolality is the Wescor Vapor Pressure Osmometer. More than 5,500 laboratories now use the Wescor VPO routinely.



Here's why it's so popular:

- Accepts any biological sample, including viscous liquids, tissue specimens and cell suspensions with no need to alter the physical state of the specimen.
- Accepts sample volumes as small as 2 microliters.
- Avoids measurement artifacts that often accompany freezing point measurements.
- Electronic accuracy and reliability without mechanical complexity.

If you are working with living cells or have other applications for accurate concentration measurements, investigate the Wescor VPO. It's the ideal osmometer.

Contact Wescor, Inc. 459 South Main Street, Logan, UT 84321 USA. (801) 752-6011 or (800) 453-2725. FAX (801) 752-4127.

WESCOR®

Innovative instrumentation since 1970.

Circle No. 33 on Readers' Service Card